

APPLICATION FOR PATENT

Inventor: Boris Filanovsky

Title: ELECTROCHEMICAL METHOD AND SENSOR FOR THE DETECTION
OF TRACES OF EXPLOSIVES

5

FIELD OF THE INVENTION

The present invention relates to the detection of trace amounts of explosive materials such as nitro-aromatic compounds in air, using an electrochemical measurement technique, and specifically to improving the sensitivity of the measurement of trace explosive materials, and decreasing measurement time. . More specifically, the present invention relates to a method of explosives detection of low cost that is suitable for portable field-testing.

BACKGROUND OF THE INVENTION

As a consequence of recent efforts to thwart the recent upsurge in international terrorism, there is an increased interest in the detection of explosive materials. These materials include nitro-aromatic compounds including 2,4,6-trinitrotoluene (TNT), dinitrotoluene (DNT) and similar derivatives.

Many detection methods have been used to detect explosive materials. These methods include gas and HPLC chromatography, x-ray scattering, neutron analysis, nuclear quadrupole resonance, and mass spectrometry (US patent 6,571,649). These methods generally require expensive and sophisticated equipment, (*e.g.* high vacuum), equipment that is not portable (*e.g.* cylinders of compressed gases), and/or have a complicated sample preparation. These techniques, are therefore, not appropriate for low cost portable field-testing for trace explosive materials. A recent review of some of these methods for explosives detection is "Explosives detection systems (EDS) for aviation security" (Singh, S., Signal Processing vol. 83, 2003, p. 31-55).

Another known method for the detection of trace amounts of explosive materials utilizes immunochemical sensors. For example, US patent 6,573,107 is directed towards the immunochemical detection of explosive substances in the gas phase using surface plasmon resonance spectroscopy. Immunochemical detection methods potentially offer high selectivity and high sensitivity.

Electrochemical detection refers to the use of electrodes, immersed in an electrolyte, and connected to an instrument that varies the voltage applied to the electrodes. The instrument measures the current flow between the electrodes. Typically, the electrode potential is varied; and an electric current flows between the electrodes that is characteristic of the presence of electrochemical active substances in the electrolyte. The magnitude of the current is proportional to the concentration of the electrochemical active substances. It is well known that TNT and other nitro-aromatic compounds are reduced electrochemically at the cathode and may be detected by electrochemical detection. Wang et. al. (*Analytica Chimica Acta*, vol. 485 (2003) p. 139-144) reported the monitoring of TNT in natural waters using an electrochemical technique. They reported a measurement sensitivity of $0.003 \mu\text{A/ppb}$ of TNT in natural seawater. This sensitivity level was achieved by Wang et. al. by subtracting the background signal, in natural seawater not contaminated by TNT, caused by the reduction of dissolved oxygen. The applicant reported (*Reviews Analytical Chemistry* vol. 18 no. 5, 1999, p. 293) the use of carbon/Hg film electrode materials in an aqueous solvent. This electrode material was successful to minimize the background by separating the atmospheric O_2 background current from the TNT current, however the sensitivity reported was only $\sim 0.7 \mu\text{A}/\mu\text{M}$ ($\sim 0.003 \mu\text{A/ppb}$) and was comparable to the sensitivity reported by Wang. Despite these positive developments in the prior art, the sensitivity is still insufficient, and the kinetics of the TNT reduction reaction

are too slow to achieve a practical portable field test for trace explosive materials. A practical electrochemical sensor for trace explosive materials should have high sensitivity, a short measurement time and in addition a way of cleaning the electrodes rapidly to perform further testing.

5 There is thus a widely recognized need for an electrochemical method and sensor for the detection of traces of explosives, and it would be highly advantageous to have an electrochemical method and sensor for the detection of traces of explosives, with high sensitivity, and fast reaction kinetics.

SUMMARY OF THE INVENTION

10 According to the present invention, there is provided a system for electrochemical assay of nitro-aromatic compounds, including: (a) a working electrode having a surface modified with a chemical that increases electron transfer kinetics of the nitro-aromatic compounds.

15 Preferably, the chemical that increases the electron transfer kinetics is an aromatic compound, for example an amino-aromatic compound, an alkyl-aniline compound, a halide derivative of an alkyl aniline compound and/or an hydroxyl-aniline compound. Most preferably, the chemical modifier is phenylene-diamine, diphenylene-diamine, diphenylene-triamine, or aniline.

20 Preferably, the working electrode contains elemental carbon or gold; the working electrode includes submicron particles and the elemental gold is a coating on the electrode surface. Preferably, the working electrode includes carbon paper.

 Preferably, the system includes in addition, (b) an electrolyte for dissolving the nitro-aromatic compounds, and the electrolyte is chosen to minimize background current resulting from oxygen reduction.

Preferably, the electrolyte includes an aprotic solvents, and/or dipolar solvents; such as dimethylformamide, acetonitrile, propylene carbonate and optionally also a solvent such as ethanol, propanol, ethylene-glycol, and/or propylene-glycol.

Preferably, the system further includes (c) a mechanism for inputting air
5 suspected to include the nitro-aromatic compounds, into the electrolyte in order to dissolve the nitro-aromatic compounds in the electrolyte.

According to the present invention, there is provided an electrochemical method of assaying trace compounds in air, including the steps of (a) dissolving the trace compounds in an electrolyte that includes aprotic solvents, and/or dipolar
10 solvents; (b) immersing a working electrode in the electrolyte; (c) applying a varying potential to the working-electrode; and (d) measuring an electrical current consequent to the varying potential, thereby providing measurement results indicative of a concentration of the trace compounds.

Preferably, after measurement, the electrochemical method includes (e)
15 regenerating the working electrode by applying a negative potential to the working electrode.

Preferably, the dissolving of trace compounds is performed by bubbling air containing the trace compounds through the electrolyte.

Preferably, the electrochemical method includes, prior to dissolving the trace
20 compounds in the electrolyte, the steps of: (f) measuring a background electrical current, while applying a varying potential, thereby obtaining background current results; and (g) subtracting the background current results from the measurement results, thereby obtaining calibrated measurement results.

Preferably, the electrolyte used in the electrochemical method includes
25 dimethylformamide, acetonitrile, and/or propylene carbonate; optionally also ethanol,

propanol, ethylene-glycol, and/or propylene-glycol and preferably the electrolyte has pH greater than 7. Preferably, the electrochemical method includes preconditioning the working electrode thereby increasing electron transfer kinetics of the trace compounds. The preconditioning modifies the working electrode with a chemical such as amino-aromatic compounds, alkyl-aniline compounds, halide derivatives of alkyl aniline compounds and/or hydroxyl- aniline compounds. According to the present invention, there is provided an electrochemical method of assaying nitro-aromatic compounds in air, including the steps of: (a) dissolving the nitro-aromatic compounds in an electrolyte that includes an aprotic solvent, and/or a dipolar solvent; (b) immersing a working electrode in the electrolyte; (c) applying a varying potential to the working electrode; (d) measuring an electrical current consequent to the varying potential, thereby providing measurement results, indicative of a concentration of the nitro-aromatic compounds.

BRIEF DESCRIPTION OF THE DRAWINGS

The invention is herein described, by way of example only, with reference to the accompanying drawings, wherein:

Fig. 1 is a drawing of an electrochemical cell used to sense trace explosive materials, according to the present invention;

Fig. 2 is a graph of background current measured in an electrochemical cell using both a conventional aqueous solvent and an electrolyte of the present invention;

Fig. 3 is a graph showing the sensitivity of the TNT detection, using an electrolyte of the present invention, and a conventional working electrode.

Fig. 4 is a graph showing the measurement sensitivity at pH=4 with a modified electrode, according to the present invention;

Fig. 5 is a graph showing the measurement sensitivity with a modified electrode and with electrolyte of pH =9, according to the present invention;

5 Fig. 6 is a graph showing the measurement sensitivity with an electrode surface further modified according to the present invention.

DESCRIPTION OF THE PREFERRED EMBODIMENTS

The present invention is an electrochemical method and sensor for the
10 detection of traces of explosives. Specifically, the present invention can be used for the detection of trace amounts in air of nitro-aromatic compounds including 2,4,6-trinitrotoluene (TNT), dinitrotoluene (DNT) and similar derivatives.

The principles and operation of an electrochemical method and sensor for the detection of traces of explosives, according to the present invention, may be better
15 understood with reference to the drawings and the accompanying description.

Referring now to the drawings, Figure 1 illustrates an electrochemical cell **101** including a working electrode **103**, a reference electrode **105**, an auxiliary electrode **106**, an electrolyte **109**, an air inlet **107** that passes air into the cell through a perforated tube **108** allowing air to bubble through electrolyte **109**, and an air outlet
20 **115** that lets air out of the cell. Reference electrode **105** is an Hg/HgCl electrode that includes an element **111** to protect reference electrode **105** from air bubbles. Electrolyte **109** is a solvent or a mixture of solvents including trace materials dissolved in the solvent(s). These trace materials, including nitro-aromatic compounds dissolved in the solvent(s), are admitted to the solvent mixture from the
25 air using air inlet **107** and dissolved in electrolyte **109** by bubbling the air through

electrolyte **109**. Air is output through air outlet **115**. A metal screen **113** is used to prevent electrolyte **109** from escaping through air outlet **115**. Voltage is applied between working electrode **103** and reference electrode **105**. A current is measured which flows between reference electrode **103** and auxiliary electrode **106**, as a result of oxidation-reduction reactions on the electrode surfaces in electrolyte **109**. Working electrode **103** is prepared by a technique of galvanic Au plating of carbon on an ordinary carbon paper surface in an aqueous solution of HAuCl , $\text{K}_4\text{Fe}(\text{CN})_6$, and Na_2CO_3 at current density $1\text{mA}/\text{cm}^2$. The carbon particles are of typical dimension $0.1\text{-}1\mu\text{m}$. The carbon paper with density $0.4\text{-}0.8\text{g}/\text{cc}$, part number P2 or P3 was supplied by E-TEK Inc. (Somerset, NJ, USA). The gold layer deposited on the carbon particles is of approximate thickness, $0.30\text{-}0.60\text{nm}$. All reagents used were obtained from Sigma-Aldrich (USA).

Voltammetric measurements were performed using a CV-50W potentiostat of Bioanalytical Systems Inc. (West Lafayette, IN, USA). Measurements were performed in the differential pulse (DIP) mode. A background current was measured using background electrolyte **109** before trace elements are introduced into electrolyte **109**. Trace elements are then introduced into electrolyte **109** through air inlet **107** using a standard air pump (not shown in Fig. 1) with throughput $1500\text{ml}/\text{min}$ for 20 sec. The current measurement is then performed and the background current is subtracted from the measured current to yield the measured results. The potential range is from -150mV to -500mV ; the scan rate is from 30 to $50\text{mV}/\text{sec}$. Electrochemical active substances in the air including nitro-aromatic compounds (*e.g.* TNT) are dissolved in electrolyte **109** and are detected electrochemically.

Figure 2 is a graph of DIP voltammetric data measured with C/Au working electrode **103**. For data trace **201**, electrolyte **109** is 0.1M KClO_4 in water with pH 9.0., For data trace **203**, electrolyte **109** is 0.1M KClO_4 in a mixture of water, ethanol, and acetonitrile (1:1:1 v/v), with pH 9.0. Figure 2 shows a window of about 250 mV, between potentials -0.5 and -0.6V where trace **302** has a lower background current. The higher background current of trace **201** is attributed to dissolved gaseous oxygen that interferes with the measurement. Consequently, the mixed electrolyte of trace **203** is preferred.

Figure 3 is a graph of DIP voltammetric data measured using an electrochemical cell shown in Figure 1, according to the present invention. However, the data of Figure 3 were measured with a working electrode C/Au **103** that was immersed in pure dimethylsulfoxide (DMSO) for 5 minutes at room temperature. Trace **301** is the background voltammetric data with electrolyte **109** of pH 4 consisting of water/ethanol/acetonitrile 1:1:1 (v/v). Trace **302** shows the voltammetric data with 200ug/l(200ppb) TNT dissolved in electrolyte **109**. Each subsequent trace **302** to **304** shows voltammetric data each with an additional 200ug/l (200ppb) TNT dissolved in electrolyte **109**. The TNT measurement sensitivity is 0.003 $\mu\text{A/ppb}$.

Figure 4 is a graph of DIP voltammetric data measured in accordance with the present invention. Working electrode **103** is preconditioned by immersing working electrode **103** in a 2% solution of aniline in DMSO for 5 minutes at room temperature. Electrolyte **109** is the same as that of Figure 3, of pH 4, consisting of water/ethanol/acetonitrile 1:1:1 (v/v). Trace **401** shows the background voltammetric data. Each subsequent trace **402** to **407** shows voltammetric data each with an

additional 60 $\mu\text{g/l}$ (60ppb) TNT dissolved in electrolyte **109**. The sensitivity of the measurement is shown in the graph of inset **410**, in which the abscissa is the TNT concentration (ppb) and the ordinate is the measured peak current in μA . The measured sensitivity is 0.11 $\mu\text{A/ppb}$. These data compared with the data of Figure 3 show that the surface modification of working electrode **103** with aniline significantly improves the measurement sensitivity of TNT.

Figure 5 is a graph of DIP voltammetric data measured in accordance with the present invention. The surface of working electrode **103** is immersed as in the measurement of Figure 4, in a 2% solution of aniline in DMSO. Electrolyte **109** is similar to that used in the measurement of Figure 4, consisting of water/ethanol/acetonitrile 1:1:1 (v/v), but with pH 9. The pH was achieved using Merck buffer capsules, no. CPM90L4 pH=9. Traces **501** to **505** show the calibrated data after subtracting the background data. Traces **501** to **505** show the calibrated data each with an additional 10 $\mu\text{g/l}$ (10 ppb) dissolved in electrolyte **109**, starting with 10 $\mu\text{g/l}$ in trace **501**. Graph inset **510** shows a sensitivity of 0.50 $\mu\text{A/ppb}$. These data compared with the measurement of Figure 4 show that the high pH increases the measurement sensitivity.

Figure 6 is a graph of DIP voltammetric data measured in accordance with the present invention. The surface of working electrode **103** is preconditioned by immersing working electrode **103** in a 3% solution of aniline in DMSO for 5 minutes at room temperature. Electrolyte **109** is the same as that of the measurement of Figure 5, consisting of water/ethanol/acetonitrile 1:1:1 (v/v), with pH 9. Trace **601** shows the background data. Traces **602** to **606** show the measured data without subtracting the background data. Traces **602** to **606** show the measured data, each when an additional 15 ppb of TNT is dissolved in electrolyte **109**, starting with 15 ppb of TNT in trace

602. The measured sensitivity is shown in graph inset 610. The measured sensitivity is 0.66 $\mu\text{A/ppb}$, showing that the stronger aniline treatment significantly improves the sensitivity.

The measurement of nitro-aromatic compounds, according to the present invention requires about 30-40 sec. After the detection of nitro-aromatic compounds, working electrode 103 is regenerated, if necessary, using a high negative potential about -1000 to -1300mV. Regeneration time is about 5 seconds. Therefore, the electrochemical technique, according to the present invention, is suitable for rapid and portable field- testing of nitro-aromatic compounds. The experimental results show that the use of solvents which are aprotic, organic dipolar, or have a high dielectric constant, as part of electrolyte 109, are expected to improve the detection sensitivity, according to the present invention. These solvents include acetonitrile, dimethylformamide and propylene- carbonate and mixtures thereof. Other polar solvents such as ethanol, propanol, ethylene-glycol, and propylene-glycol may be added as diluents. A preferred diluent has a high boiling point and is stable against evaporation. The electrode used, according to the present invention, is manufactured from ordinary carbon paper. Carbon paper is readily available, of low cost and therefore suitable for a disposable electrode. An electrode, including a carbon particle layer on other substrates including cloth or glass, will function in a similar way.

The experimental results furthermore show that the chemical modification of a, carbon or carbon/gold electrode, with compounds similar to aniline, preconditions the electrode surface to increase the sensitivity of the measurement, according to the present invention. These compounds include aromatic compounds containing amino groups, , including derivatives that are mono-alkyl (*e.g.* methyl, ethyl, propyl,...), di-alkyl or tri-alkyl. The chemical modifiers, according to the present invention include

aromatic-aniline compounds such as, phenylene-diamine, diphenylene-diamine, diphenylene-triamine and similar compounds and derivatives. This chemical modification is necessary because the reduction of nitro-aromatic compounds has slow kinetics. The modification successfully increases the electron transfer rate from
5 the solution to electrode **103** and therefore increases the sensitivity of the measurement and lowers the detection limit. Other modifiers that increase electron transfer, according to the present invention, include compounds such as J. Meisenheimers complexes, known to mediate electron transfer, and other nitro-amine complexes; and alkyl-aniline and its derivatives; and halide derivatives of alkyl aniline
10 compounds as well as hydroxyl- aniline compounds.

While the invention has been described with respect to a limited number of embodiments, it will be appreciated that many variations, modifications and other applications of the invention may be made.